

MAR 29 2011

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130300</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>JEREMY FERNANDES</u>	ID# <u>6345</u> ✓ Date <u>3-9-11</u> ✓
A Agency <u>FYPD</u>	Phone # <u>662-2311</u>
Instrument Location <u>FOTET YUKON</u>	
B Alco S/N <u>X124617</u> ✓ Target Value <u>.084</u> ✓ High Pressure <u>400</u>	
Alco Test Values	
1st Alco <u>.088</u> ✓	2nd Alco <u>.087</u> ✓
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 1<sup>ST</sup> day of APRIL, 2011

[Signature]  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130300 *bmb 3/31/11*

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130300

MARCH 09, 2011 *bmb 3/31/11*

OPERATOR'S NAME:

FERNANDES/JEREMY/GILMORE

OPERATOR'S NUMBER: 6345

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :

OF CAL

O.L. #: 00000

DEPT/AGENCY: FYU1

CASE/REPORT: 11

TEST TYPE: T

ALCO TARGET VALUE: .004

ALCO S/N: X124617

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 30.06 in  
ALCO TARGET .004 13:23  
BLANK TEST .000 13:24  
INTERNAL STANDARD VERIFIED 13:24  
ALCO TV 30.06 in .000 13:24  
BLANK TEST .000 13:25  
SUBJECT SAMPLE *bmb 3/31/11* .000 13:26  
BLANK TEST .000 13:27  
ALCO TV 30.06 in .007 13:27  
BLANK TEST .000 13:28

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130300

MARCH 09, 2011

TIME 13:16

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 30.06 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; ( = ) ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢